

Digital Hydraulic Pinch Gauge

Instructions for Use



REF 12-0237

Intended Purpose

The digital hydraulic pinch gauge is a medical/therapeutic device designed to measure finger pinch strength during tip, key, and palmar pinch testing.

The finger pinch gauge can be used to measure pinch strength. It is calibrated in pounds and kilograms of force.

USE: Reset max indicator to zero. Apply pinch force at the pinch groove while holding the pinch gauge between your thumb and finger(s). When force is applied farther toward the tip the reading will be slightly higher. When force is applied farther toward the rear the reading will be slightly lower.

The gauge must be “zeroed” before each pinch test. Refer to instructions on page two for how to use digital settings.

Use the pinch gauge to perform the three basic pinch tests:

- **Tip Pinch** - thumb tip to index fingertip
- **Key Pinch** -thumb pad to lateral aspect of middle phalanx of index finger
- **Palmar Pinch** - thumb pad to pads of the index and middle fingers


Digital Setting Instructions

How to turn on/off unit:

Press

ON
ENTER

- Press ON/ENTER to switch ON and OFF.



How to set max indicator to zero:

Hold

MENU

Zero Screen

ZERO
UNIT

Hold

ON
ENTER

- Display "----" on the LCD. Zero the unit and save the Zero Offset.

How to view max:

Hold

MENU

Zero Screen

ZERO
UNIT

Press

MENU

Max Screen

MAX

Hold

ON
ENTER

- Exit without changing anything.

How to clear max:

Hold

MENU

Zero Screen

ZERO
UNIT

Press

MENU

Max Screen

MAX

Press

MENU

Max Clear
Screen

MAX
CLR

Hold

ON
ENTER

Confirm
Clear

CLR
YES?

Hold

ON
ENTER

- Display "----" on the LCD. Clear MAX to 0 and save the MAX value.

How to switch between LB and KG:

Hold

MENU

Zero Screen

ZERO
UNIT

Press

MENU

Max Screen

MAX

Press

MENU

Max Clear
Screen

MAX
CLR

Press

MENU

Select Unit
Screen

LBKG
SEL

Hold

ON
ENTER

Press

MENU

Hold

ON
ENTER

- Switch the weight unit (LB and KG)
- Display "----" on the LCD. Save the weight unit.

Parts/Specifications



Specifications: 100 lb / 45 kg capacity

Tolerances:

The digital 100 lb. hydraulic pinch gauge is tested for accuracy at 15, 30, 40, 50, 60, 70, 85 and 100 lbs. TOLERANCES (Based on 2% through entire range).

Disposal Method

Dispose of item in accordance with the local/regional/national/international regulations.



WARNING:

- Clean contact surfaces between each patient
- Verify zero position before measurements
- Ensure proper finger placement
- Do not exceed device's force range
- Stop if patient experiences discomfort
- Keep device properly calibrated
- Store in protective case
- Inspect for loose parts before use
- Check gauge accuracy regularly

norms for adult GRIP strength performance of all subjects (lbs)

age	hand	men			women		
		mean	SD	low-high	mean	SD	low-high
20-24	dominant	121.0	20.6	91-167	70.4	14.5	46-95
	non-dominant	104.5	21.8	71-150	61.0	13.1	33-88
25-29	dominant	120.8	23.0	78-158	74.5	13.9	48-97
	non-dominant	110.5	16.2	77-139	63.5	12.2	48-97
30-34	dominant	121.8	22.4	70-170	78.7	19.2	46-137
	non-dominant	110.4	21.7	64-145	68.0	17.7	36-115
35-39	dominant	119.7	24.0	76-176	74.1	10.8	50-99
	non-dominant	112.9	21.7	73-157	66.3	11.7	49-91
40-44	dominant	116.8	20.7	84-165	70.4	13.5	38-103
	non-dominant	112.8	18.7	73-157	62.3	13.8	35-94
45-49	dominant	109.9	23.0	65-155	62.2	15.1	39-100
	non-dominant	100.8	22.8	58-160	56.0	12.7	37-83
50-54	dominant	113.6	18.1	79-151	65.8	11.6	38-87
	non-dominant	101.9	17.0	70-143	57.3	10.7	35-76
55-59	dominant	101.1	26.7	59-154	57.3	12.5	33-86
	non-dominant	83.2	23.4	43-128	47.3	11.9	31-76
60-64	dominant	89.7	20.4	51-137	55.1	10.1	37-77
	non-dominant	76.8	20.3	27-116	45.7	10.1	29-66
65-69	dominant	91.1	20.6	56-131	49.6	9.7	35-74
	non-dominant	76.8	19.8	43-117	41.0	8.2	29-63
70-75	dominant	75.3	21.5	32-108	49.6	11.7	33-78
	non-dominant	64.8	18.1	32-93	41.5	10.2	23-67
75+	dominant	65.7	21.1	40-135	42.6	11.0	25-65
	non-dominant	55.0	17.0	31-119	37.6	8.9	24-61
ALL	dominant	104.3	28.3	32-176	62.8	17.0	25-137
	non-dominant	93.1	27.6	27-160	53.9	15.7	23-115

norms for adult PINCH strength (Tip Pinch strength) performance of all subjects (lbs)

age	hand	men			women		
		mean	SD	low-high	mean	SD	low-high
20-24	dominant	18.0	3.0	11-23	11.1	2.1	8-16
	non-dominant	17.0	2.3	12-33	10.5	1.7	8-14
25-29	dominant	18.3	4.4	10-34	11.9	1.8	8-16
	non-dominant	17.5	5.2	12-36	11.3	1.8	9-18
30-34	dominant	17.4	6.7	12-25	12.6	3.0	8-20
	non-dominant	17.6	4.8	10-27	11.7	2.8	7-17
35-39	dominant	18.0	3.6	12-27	11.6	2.5	8-19
	non-dominant	17.7	3.8	10-24	11.9	2.4	8-16
40-44	dominant	17.8	4.0	11-25	11.5	2.7	5-15
	non-dominant	17.7	3.5	12-25	11.1	3.0	6-17
45-49	dominant	18.7	4.9	12-30	13.2	3.0	9-19
	non-dominant	17.6	4.1	12-28	12.1	2.7	7-18
50-54	dominant	18.3	4.0	11-24	12.5	2.2	9-18
	non-dominant	17.8	3.9	12-26	11.4	2.4	7-16
55-59	dominant	16.6	3.3	11-24	11.7	1.7	9-16
	non-dominant	15.0	3.7	10-26	10.4	1.4	8-13
60-64	dominant	15.8	3.9	9-22	10.1	2.1	7-17
	non-dominant	15.3	3.7	9-23	9.9	2.0	6-15
65-69	dominant	17.0	4.2	11-27	10.6	2.0	7-15
	non-dominant	15.4	2.9	10-21	10.5	2.4	7-17
70-75	dominant	13.8	2.6	11-21	10.1	2.6	7-15
	non-dominant	13.3	2.6	10-21	9.8	2.3	6-17
75+	dominant	14.0	3.4	7-21	9.6	2.8	4-16
	non-dominant	13.9	3.7	8-25	9.3	2.4	4-13
ALL	dominant						
	non-dominant						

norms for adult PINCH strength (Key Pinch strength)
performance of all subjects (lbs)

age	hand	men		low-high	women		low-high
		mean	SD		mean	SD	
20-24	dominant	26.0	3.5	21-34	17.6	2.0	14-23
	non-dominant	24.8	3.4	19-31	16.2	2.1	13-23
25-29	dominant	26.7	4.9	19-41	17.7	2.1	14-22
	non-dominant	25.0	4.7	19-39	16.6	2.1	13-22
30-34	dominant	26.4	4.8	20-36	18.7	3.0	13-25
	non-dominant	26.2	5.1	17-36	17.8	3.6	12-26
35-39	dominant	26.1	3.2	21-32	16.6	2.0	12-21
	non-dominant	25.6	3.9	18-32	16.0	2.7	12-22
40-44	dominant	25.6	2.6	21-31	16.7	3.1	10-24
	non-dominant	25.1	4.0	19-31	15.8	3.1	8-22
45-49	dominant	25.8	3.9	19-35	17.6	3.2	13-24
	non-dominant	24.8	4.4	18-42	16.6	2.9	12-24
50-54	dominant	26.7	4.4	20-34	16.7	2.5	12-22
	non-dominant	26.1	4.2	20-37	16.1	2.7	12-22
55-59	dominant	24.2	4.2	18-34	15.7	2.5	11-21
	non-dominant	23.0	4.7	13-31	14.7	2.2	12-19
60-64	dominant	23.2	5.4	14-37	15.5	2.7	10-20
	non-dominant	22.2	4.1	16-33	14.1	2.5	10-19
65-69	dominant	23.4	3.9	17-32	15.0	2.6	10-21
	non-dominant	22.0	3.6	17-28	14.3	2.8	10-20
70-75	dominant	19.3	2.4	16-25	14.5	2.9	8-22
	non-dominant	19.2	3.0	13-28	13.8	3.0	9-22
75+	dominant	20.5	4.6	9-31	12.6	2.3	8-17
	non-dominant	19.1	3.0	13-24	11.4	2.6	7-16
ALL	dominant	24.5	4.6	9-41	16.2	3.0	8-25
	non-dominant	23.6	4.6	11-42	15.3	3.1	7-26

norms for adult PINCH strength (Palmar Pinch strength)
performance of all subjects (lbs)

age	hand	men		low-high	women		low-high
		mean	SD		mean	SD	
20-24	dominant	26.6	5.3	18-45	17.2	2.3	14-23
	non-dominant	25.7	5.8	15-42	16.3	2.8	11-24
25-29	dominant	26.0	4.3	19-35	17.7	3.2	13-29
	non-dominant	25.1	4.2	19-36	17.0	3.0	13-26
30-34	dominant	24.7	4.7	16-34	19.3	5.0	12-34
	non-dominant	25.4	5.7	15-37	18.1	4.8	12-32
35-39	dominant	26.2	4.1	19-36	17.5	4.2	13-29
	non-dominant	25.9	5.4	14-40	17.1	3.4	12-24
40-44	dominant	24.5	4.3	17-37	17.0	3.1	10-23
	non-dominant	24.8	4.9	15-37	16.6	3.5	14-25
45-49	dominant	24.0	3.3	19-33	17.9	3.0	12-27
	non-dominant	23.7	3.8	8-33	17.5	2.8	12-24
50-54	dominant	23.8	5.4	15-36	17.3	3.1	12-23
	non-dominant	24.0	5.8	16-36	16.4	2.9	12-22
55-59	dominant	23.7	4.8	16-34	16.0	3.1	11-26
	non-dominant	21.3	4.5	12-25	15.4	3.0	11-21
60-64	dominant	21.8	3.3	16-28	14.8	3.1	10-20
	non-dominant	21.2	3.2	15-27	14.3	2.7	10-20
65-69	dominant	21.4	3.0	15-25	14.2	3.1	8-20
	non-dominant	21.2	4.1	14-30	13.7	3.4	8-22
70-75	dominant	18.1	3.4	14-27	14.4	2.6	9-19
	non-dominant	18.8	3.3	13-27	14.0	1.9	10-17
75+	dominant	18.7	4.2	9-26	12.0	2.6	8-17
	non-dominant	18.3	3.8	10-26	11.5	2.6	6-16
ALL	dominant	23.4	5.0	9-45	16.3	3.8	8-34
	non-dominant	23.0	5.3	10-42	15.7	3.6	6-32

 Fabrication Enterprises Inc
250 Clearbrook Road, Suite 240
Elmsford, NY 10523 (USA)
tel: +1-914-345-9900 • 800-431-2830
fax: +1-914-345-9800 • 800-634-5370
FabEnt.com



AJW Technology Consulting GmbH
Breite Strasse 3
40213 Dusseldorf Germany



MedEnvoy
NL-IM-000000248
Prinses Margrietplantsoen 33, Suite 123
2595 AM, The Hague (NL)